
Midmark M11 - v1.0.5
Total Cycles: 3052

Name: RITTER2
Sterilizer ID: V2313228
10/25/2023 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 158.2 F 1.7
4:00 171.9 F 5.0
6:00 193.7 F 9.5
8:00 220.9 F 7.8
10:00 226.1 F 9.3
12:00 234.7 F 9.8
14:00 241.1 F 11.9
16:00 239.7 F 10.5
18:00 252.8 F 17.1
20:00 264.3 F 23.9
22:00 270.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.0 F 29.3
2:00 272.7 F 29.6
3:00 272.7 F 29.9
4:00 271.5 F 29.6

Min 271.3 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 238.0 F 0.0
10:00 186.7 F 0.0
15:00 180.8 F 0.0
20:00 179.6 F 0.0
25:00 178.5 F 0.0
30:00 178.2 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:26
CYCLE COMPLETE

10/25/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____