
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001716
10:16 09/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:49
STERILIZING 3:53
136.0 C 220kPa 3:53
136.2 C 223kPa 4:23
136.2 C 222kPa 4:53
136.2 C 223kPa 5:23
136.0 C 221kPa 5:53
136.3 C 222kPa 6:23
136.3 C 223kPa 6:53
136.4 C 223kPa 7:23
136.4 C 223kPa 7:23
VENTING 7:23
CYCLE ABORTED 7:45
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:45

Digital Signature #
4C1D00295EF7FB2F
