
Midmark M11 - v1.0.0
Total Cycles: 4964

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2024 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 77.9 F 0.0
4:00 89.5 F 0.0
6:00 118.8 F 0.3
8:00 143.9 F 1.5
10:00 173.6 F 4.2
12:00 191.7 F 7.2
14:00 205.1 F 11.2
16:00 215.0 F 3.8
18:00 224.6 F 7.5
20:00 235.0 F 11.7
22:00 232.7 F 7.7
24:00 245.3 F 13.3
26:00 242.8 F 11.5
28:00 254.8 F 17.7
30:00 265.5 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 27.9
3:00 272.7 F 28.9
4:00 272.0 F 28.2

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 192.5 F 0.0
10:00 198.6 F 0.0
15:00 201.4 F 0.0
20:00 198.5 F 0.0
25:00 199.3 F 0.0
30:00 206.9 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

10/04/2024 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____