
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001500
8:52 10/14/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:38
PRESSURIZING 6:20
STERILIZING 8:10
136.0 C 220kPa 8:10
136.5 C 221kPa 8:40
136.3 C 219kPa 9:10
136.4 C 221kPa 9:40
136.4 C 222kPa 10:10
136.2 C 221kPa 10:40
136.2 C 221kPa 11:10
136.4 C 221kPa 11:40
VENTING 11:40
AIR DRYING 12:47
CYCLE COMPLETE 42:30

Digital Signature #
66E88FADE0B6B47E
