

-----  
Midmark M11 - v1.0.0  
Total Cycles: 959

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2021 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 214.1 F 0.0  
2:00 193.8 F 3.4  
4:00 216.8 F 8.0  
6:00 220.0 F 13.5  
8:00 228.5 F 8.2  
10:00 241.2 F 13.4  
12:00 241.2 F 11.1  
14:00 238.4 F 9.4  
16:00 252.0 F 16.1  
18:00 263.4 F 22.6  
20:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.8 F 28.8  
2:00 272.8 F 28.9  
3:00 272.5 F 29.0  
4:00 271.6 F 28.7

Min 271.0 F 28.0  
Max 273.0 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 221.3 F 0.0  
10:00 212.3 F 0.0  
15:00 210.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:50  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 19:40  
TOTAL CYCLE: 00:48:51

07/05/2021 10:56 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01