
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000551
11:33 23/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:24
STERILIZING 4:41
136.0 C 220kPa 4:41
136.3 C 233kPa 5:11
136.4 C 232kPa 5:41
136.3 C 232kPa 6:11
136.4 C 233kPa 6:41
136.2 C 231kPa 7:11
136.3 C 231kPa 7:41
136.3 C 231kPa 8:11
136.3 C 231kPa 8:11
VENTING 8:11
AIR DRYING 9:06
CYCLE COMPLETE 24:07

Digital Signature #
1E678408E35F3658
