

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1550

Name: RITTER2  
Sterilizer ID: V2313228  
10/03/2022 08:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
2:00 164.3 F 0.0  
4:00 173.6 F 0.3  
6:00 199.0 F 4.2  
8:00 228.3 F 9.6  
10:00 229.7 F 8.0  
12:00 239.1 F 12.2  
14:00 238.9 F 10.6  
16:00 238.1 F 9.5  
18:00 252.7 F 16.8  
20:00 265.0 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.6 F 29.3  
2:00 271.5 F 28.4  
3:00 272.5 F 29.4  
4:00 272.3 F 28.8  
  
Min 271.2 F 28.4  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 227.3 F 0.0  
10:00 247.1 F 0.0  
15:00 228.9 F 0.0  
20:00 227.2 F 0.0  
25:00 231.4 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 21:50  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

10/03/2022 09:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_