

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002345  
11:56 03/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:45  
STERILIZING 3:57  
136.2 C 223kPa 3:57  
136.3 C 225kPa 4:27  
136.3 C 225kPa 4:57  
136.4 C 225kPa 5:27  
136.4 C 225kPa 5:57  
136.3 C 224kPa 6:27  
136.2 C 223kPa 6:57  
136.3 C 225kPa 7:27  
136.3 C 225kPa 7:27  
VENTING 7:27  
AIR DRYING 8:26  
CYCLE COMPLETE 8:37  
-----

Digital Signature #  
178B222DB4E10904  
-----