
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002597
8:35 02/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:49
STERILIZING 4:54
136.5 C 224kPa 4:54
136.3 C 224kPa 5:24
136.4 C 224kPa 5:54
136.4 C 224kPa 6:24
136.3 C 223kPa 6:54
136.4 C 221kPa 7:24
136.4 C 223kPa 7:54
136.4 C 220kPa 8:24
136.4 C 220kPa 8:24
VENTING 8:24
AIR DRYING 9:20
CYCLE COMPLETE 9:22

Digital Signature #
02DE2EF4DD690A9E
