
Midmark M11 - v1.0.5
Total Cycles: 20

Name: RITTER2
Sterilizer ID: V2313228
04/26/2021 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 182.5 F 3.7
4:00 203.0 F 9.1
6:00 224.3 F 7.5
8:00 233.2 F 11.3
10:00 235.9 F 9.6
12:00 237.3 F 9.1
14:00 251.0 F 16.3
16:00 263.9 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.4 F 29.3
2:00 272.6 F 29.8
3:00 271.7 F 29.1
4:00 272.5 F 29.6

Min 271.6 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.7 F 0.0
5:00 217.4 F 0.0
10:00 209.1 F 0.0
15:00 209.0 F 0.0
20:00 209.6 F 0.0
25:00 183.9 F 0.0
30:00 174.7 F 0.0

Drying Complete

FILLING: 2:26
HEATING: 17:54
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:28
CYCLE COMPLETE

04/26/2021 02:40 PM
ACCEPTED: X
REJECTED:
Operator: U01