

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2600

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2022 08:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.5 F 0.0  
2:00 85.2 F 0.0  
4:00 97.5 F 0.0  
6:00 117.8 F 0.4  
8:00 145.7 F 3.3  
10:00 167.4 F 7.0  
12:00 191.0 F 10.7  
14:00 216.7 F 5.4  
16:00 218.7 F 6.3  
18:00 229.5 F 10.2  
20:00 237.2 F 9.3  
22:00 240.8 F 11.3  
24:00 237.4 F 9.0  
26:00 250.3 F 15.3  
28:00 261.3 F 21.5  
30:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 28.8  
1:00 272.6 F 29.1  
2:00 272.4 F 29.3  
3:00 271.9 F 29.1  
4:00 272.2 F 28.7

Min 271.8 F 28.7  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 224.5 F 0.0  
10:00 227.9 F 0.0  
15:00 228.3 F 0.0  
20:00 222.9 F 0.0  
25:00 221.6 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:14  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

11/23/2022 09:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_