
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003029
12:02 26/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:29
STERILIZING 3:25
136.4 C 222kPa 3:25
136.3 C 223kPa 3:55
136.2 C 222kPa 4:25
136.4 C 222kPa 4:55
136.4 C 225kPa 5:25
136.3 C 224kPa 5:55
136.4 C 223kPa 6:25
136.1 C 220kPa 6:55
136.1 C 220kPa 6:55
VENTING 6:55
AIR DRYING 7:45
CYCLE COMPLETE 7:48

Digital Signature #
473B4B52013D0DEA
