
Midmark M11 - v1.0.5
Total Cycles: 3091

Name: RITTER2
Sterilizer ID: V2313228
11/07/2023 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.8 F 0.0
4:00 76.9 F 0.3
6:00 107.8 F 1.8
8:00 146.6 F 4.7
10:00 172.7 F 7.8
12:00 215.1 F 11.3
14:00 215.1 F 4.0
16:00 225.2 F 8.0
18:00 235.9 F 12.2
20:00 232.2 F 8.4
22:00 243.4 F 12.5
24:00 242.7 F 12.2
26:00 255.0 F 18.5
28:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.1 F 29.3
2:00 272.3 F 29.8
3:00 271.2 F 29.0
4:00 272.1 F 29.5

Min 271.2 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 231.1 F 0.0
10:00 196.2 F 0.0
15:00 192.3 F 0.0
20:00 187.6 F 0.0
25:00 185.1 F 0.0
30:00 185.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:40
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:07:42
CYCLE COMPLETE

11/07/2023 10:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____