
Midmark M11 - v1.0.0
Total Cycles: 2076

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2022 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.9 F 0.0
2:00 162.8 F 2.1
4:00 173.5 F 5.4
6:00 199.3 F 9.6
8:00 216.5 F 5.5
10:00 226.2 F 8.5
12:00 241.4 F 13.4
14:00 238.5 F 10.4
16:00 237.2 F 7.9
18:00 249.7 F 15.1
20:00 261.6 F 21.8
22:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.1
2:00 272.1 F 28.6
3:00 271.2 F 28.2
4:00 272.2 F 29.1

Min 271.2 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 215.4 F 0.0
10:00 239.4 F 0.0
15:00 237.8 F 0.0
20:00 237.9 F 0.0
25:00 226.4 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

06/29/2022 03:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____