

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1858

Name: RITTER 1  
Sterilizer ID: V2178198  
04/22/2022 01:42 PM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	170.3 F	0.0
2:00	171.3 F	1.5
4:00	185.9 F	4.7
6:00	204.2 F	9.0
8:00	219.4 F	10.8
10:00	228.4 F	7.5
12:00	240.1 F	12.4
14:00	238.3 F	9.6
16:00	238.0 F	8.7
18:00	248.9 F	14.2
20:00	252.8 F	16.2

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 251F  
CHAMBER PRESS 15.1PSI  
\*\* ITEMS NOT STERILE \*\*

04/22/2022 02:04 PM