

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1279

Name: RITTER 1  
Sterilizer ID: V2178198  
10/20/2021 06:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.7 F 0.0  
2:00 73.1 F 0.0  
4:00 81.8 F 0.0  
6:00 104.3 F 0.0  
8:00 141.1 F 0.1  
10:00 176.2 F 0.2  
12:00 195.5 F 0.2  
14:00 211.1 F 2.3  
16:00 220.1 F 4.9  
18:00 227.1 F 7.5  
20:00 235.0 F 10.7  
22:00 235.6 F 8.8  
24:00 242.3 F 11.8  
26:00 238.7 F 9.8  
28:00 238.3 F 9.1  
30:00 249.5 F 14.7  
32:00 260.5 F 20.8  
34:00 270.5 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.8 F 28.7  
2:00 271.6 F 27.8  
3:00 272.6 F 28.7  
4:00 272.1 F 28.2  
  
Min 271.4 F 27.8  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.1 F 0.0  
5:00 221.0 F 0.0  
10:00 221.5 F 0.0  
15:00 225.9 F 0.0  
20:00 236.4 F 0.0  
25:00 194.6 F 0.0  
30:00 191.8 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 34:52  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:13:06  
CYCLE COMPLETE

10/20/2021 07:51 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01