
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000547
12:43 16/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:26
STERILIZING 4:37
136.0 C 228kPa 4:37
136.4 C 230kPa 5:07
136.1 C 230kPa 5:37
136.2 C 231kPa 6:07
136.4 C 232kPa 6:37
136.2 C 230kPa 7:07
136.3 C 231kPa 7:37
136.2 C 229kPa 8:07
136.2 C 229kPa 8:07
VENTING 8:07
AIR DRYING 9:01
CYCLE COMPLETE 24:01

Digital Signature #
B0E3A7DE6ED94CED
