
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005156
14:17 27/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:07
STERILIZING 3:59
136.0 C 220kPa 3:59
136.3 C 225kPa 4:29
136.4 C 221kPa 4:59
136.3 C 225kPa 5:29
136.4 C 226kPa 5:59
136.1 C 223kPa 6:29
136.4 C 224kPa 6:59
136.4 C 223kPa 7:29
136.4 C 223kPa 7:29
VENTING 7:29
AIR DRYING 8:20
CYCLE COMPLETE 19:40

Digital Signature #
A65878B9E6D13260
