
Midmark M11 - v1.0.5
Total Cycles: 3861

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 167.6 F 2.9
4:00 188.8 F 7.0
6:00 210.4 F 11.4
8:00 220.2 F 6.4
10:00 232.3 F 11.3
12:00 232.8 F 8.7
14:00 240.0 F 10.7
16:00 245.6 F 13.7
18:00 256.6 F 19.8
20:00 267.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.5 F 30.0
2:00 271.6 F 29.4
3:00 271.4 F 29.8
4:00 272.7 F 30.8

Min 270.9 F 29.2
Max 272.9 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 208.0 F 0.0
10:00 196.8 F 0.0
15:00 194.3 F 0.0
20:00 192.1 F 0.0
25:00 189.2 F 0.0
30:00 186.7 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

08/14/2024 04:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____