
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000183
11:12 06/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:39
PRESSURIZING 4:39
STERILIZING 5:57
136.4 C 224kPa 5:57
136.3 C 224kPa 6:27
136.4 C 225kPa 6:57
136.1 C 219kPa 7:27
136.1 C 223kPa 7:57
136.0 C 221kPa 8:27
136.2 C 223kPa 8:57
136.4 C 225kPa 9:27
136.4 C 225kPa 9:27
VENTING 9:27
AIR DRYING 10:28
CYCLE COMPLETE 10:39

Digital Signature #
E68A7840C43DD18B
