
Midmark M11 - v1.0.5
Total Cycles: 3460

Name: RITTER2
Sterilizer ID: V2313228
03/20/2024 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.9 F 0.0
2:00 121.1 F 0.1
4:00 124.8 F 1.6
6:00 154.1 F 4.9
8:00 177.3 F 8.6
10:00 197.8 F 12.9
12:00 211.6 F 6.2
14:00 228.3 F 10.7
16:00 229.0 F 7.7
18:00 242.1 F 13.1
20:00 241.5 F 11.9
22:00 254.0 F 18.1
24:00 264.9 F 24.8
26:00 271.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.9
1:00 272.6 F 30.0
2:00 273.1 F 30.4
3:00 272.0 F 30.1
4:00 272.4 F 30.0

Min 271.8 F 29.7
Max 273.1 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 206.0 F 0.0
10:00 192.4 F 0.0
15:00 190.8 F 0.0
20:00 190.0 F 0.0
25:00 191.0 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:39
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:05:04
CYCLE COMPLETE

03/20/2024 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____