
Midmark M11 - v1.0.0
Total Cycles: 5151

Name: RITTER 1
Sterilizer ID: V2178198
12/03/2024 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 147.7 F 0.1
4:00 166.4 F 0.4
6:00 189.0 F 1.0
8:00 208.5 F 4.6
10:00 222.3 F 9.0
12:00 235.1 F 12.1
14:00 236.1 F 10.3
16:00 235.6 F 8.0
18:00 242.6 F 10.9
20:00 247.5 F 13.7
22:00 259.8 F 20.4
24:00 270.8 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.0
1:00 272.4 F 28.8
2:00 272.0 F 28.3
3:00 271.1 F 27.8
4:00 272.2 F 28.9

Min 271.0 F 27.8
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 217.4 F 0.0
10:00 194.2 F 0.0
15:00 194.5 F 0.0
20:00 214.2 F 0.0
25:00 215.8 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:50
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:15
CYCLE COMPLETE

12/03/2024 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____