

-----  
Midmark M11 - v1.0.0  
Total Cycles: 611

Name: MIDMARK  
Sterilizer ID: V2178198  
01/20/2021 10:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.5 F 0.0  
2:00 96.0 F 0.0  
4:00 122.6 F 0.1  
6:00 155.2 F 0.5  
8:00 183.8 F 3.9  
10:00 203.6 F 7.9  
12:00 218.1 F 12.6  
14:00 225.6 F 7.0  
16:00 238.4 F 12.1  
18:00 235.8 F 9.1  
20:00 238.7 F 9.4  
22:00 247.1 F 13.9  
24:00 259.5 F 20.6  
26:00 270.8 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.3 F 28.7  
2:00 272.6 F 29.2  
3:00 271.7 F 28.3  
4:00 272.6 F 29.0

Min 271.4 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 165.7 F 0.0  
5:00 215.1 F 0.0  
10:00 230.8 F 0.0  
15:00 224.9 F 0.0  
20:00 221.9 F 0.0  
25:00 219.9 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:43  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:04:37  
CYCLE COMPLETE

01/20/2021 11:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_