

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3974

Name: RITTER2  
Sterilizer ID: V2313228  
09/25/2024 01:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.1 F 0.0  
2:00 114.5 F 0.1  
4:00 131.7 F 0.5  
6:00 163.4 F 3.8  
8:00 181.5 F 7.4  
10:00 226.3 F 11.2  
12:00 218.7 F 5.5  
14:00 232.6 F 10.1  
16:00 231.8 F 8.6  
18:00 241.0 F 12.1  
20:00 239.6 F 10.6  
22:00 252.0 F 16.8  
24:00 263.2 F 23.2  
26:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.2 F 29.6  
2:00 271.2 F 28.8  
3:00 271.8 F 29.7  
4:00 272.0 F 29.2

Min 271.0 F 28.7  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.8 F 0.0  
5:00 226.2 F 0.0  
10:00 187.5 F 0.0  
15:00 181.9 F 0.0  
20:00 177.3 F 0.0  
25:00 158.7 F 0.0  
30:00 151.9 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:18  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:04:47  
CYCLE COMPLETE

09/25/2024 02:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_