

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5264

Name: RITTER 1  
Sterilizer ID: V2178198  
01/22/2025 08:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
2:00 181.2 F 4.2  
4:00 196.8 F 8.7  
6:00 221.2 F 9.6  
8:00 230.9 F 8.3  
10:00 243.1 F 13.5  
12:00 240.2 F 10.8  
14:00 237.3 F 9.0  
16:00 251.8 F 16.0  
18:00 263.9 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.6  
2:00 271.8 F 28.6  
3:00 272.3 F 28.4  
4:00 271.9 F 28.7  
  
Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 210.0 F 0.0  
10:00 204.2 F 0.0  
15:00 235.9 F 0.0  
20:00 238.8 F 0.0  
25:00 235.2 F 0.0  
30:00 224.2 F 0.0  
Drying Complete

FILLING: 3:44  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:01:18  
CYCLE COMPLETE

01/22/2025 09:33 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_