

-----  
Midmark M11 - v1.0.5  
Total Cycles: 259

Name: RITTER2  
Sterilizer ID: V2313228  
07/28/2021 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
2:00 185.5 F 3.2  
4:00 189.0 F 7.2  
6:00 206.0 F 11.7  
8:00 224.6 F 6.0  
10:00 235.1 F 10.6  
12:00 233.5 F 8.3  
14:00 244.1 F 13.4  
16:00 244.2 F 12.6  
18:00 255.7 F 18.7  
20:00 266.3 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.5 F 29.0  
2:00 272.3 F 29.7  
3:00 271.7 F 29.4  
4:00 271.6 F 29.1

Min 270.9 F 28.6  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
5:00 225.3 F 0.0  
10:00 217.3 F 0.0  
15:00 218.4 F 0.0  
20:00 220.4 F 0.0  
25:00 229.4 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 21:40  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

07/28/2021 01:01 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01