
Midmark M11 - v1.0.5
Total Cycles: 83

Name: RITTER2
Sterilizer ID: V2313228
06/09/2021 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.3 F 0.0
2:00 158.8 F 3.4
4:00 178.0 F 8.0
6:00 202.9 F 11.6
8:00 219.8 F 7.7
10:00 234.5 F 12.2
12:00 238.6 F 11.6
14:00 239.4 F 10.7
16:00 253.4 F 17.8
18:00 265.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.7 F 29.7
2:00 272.0 F 29.6
3:00 272.4 F 29.5
4:00 272.0 F 29.7

Min 271.7 F 29.0
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 222.6 F 0.0
10:00 229.2 F 0.0
15:00 234.1 F 0.0
Dry Cycle Aborted

FILLING: 1:12
HEATING: 19:39
STERILIZING: 4:00
VENTING: 2:53
DRYING: 15:10
TOTAL CYCLE: 00:42:54

06/09/2021 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____