
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005670
13:26 05/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:03
STERILIZING 4:08
136.5 C 220kPa 4:08
136.4 C 225kPa 4:39
136.4 C 225kPa 5:09
136.3 C 223kPa 5:39
136.3 C 223kPa 6:09
136.4 C 225kPa 6:39
136.1 C 218kPa 7:09
136.2 C 221kPa 7:39
136.2 C 221kPa 7:39
VENTING 7:39
AIR DRYING 8:37
CYCLE COMPLETE 11:50

Digital Signature #
816278CDDC9191D9
