
Midmark M11 - v1.0.0
Total Cycles: 3843

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2023 02:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.6 F 0.0
2:00 121.7 F 0.0
4:00 149.0 F 0.5
6:00 178.3 F 2.9
8:00 198.6 F 6.0
10:00 209.7 F 9.9
12:00 218.5 F 11.3
14:00 227.3 F 8.5
16:00 238.7 F 13.0
18:00 238.2 F 10.2
20:00 237.1 F 8.7
22:00 249.3 F 14.7
24:00 260.8 F 21.1
26:00 271.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.2 F 28.5
2:00 271.5 F 28.3
3:00 272.1 F 28.5
4:00 271.3 F 28.3

Min 271.2 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 205.4 F 0.0
10:00 204.6 F 0.0
15:00 195.8 F 0.0
20:00 212.5 F 0.0
25:00 235.3 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:39
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

09/29/2023 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____