

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001983  
15:05 26/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:06  
STERILIZING 3:16  
136.5 C 226kPa 3:16  
136.3 C 226kPa 3:46  
136.3 C 222kPa 4:16  
136.4 C 224kPa 4:46  
136.4 C 225kPa 5:16  
136.4 C 226kPa 5:46  
136.2 C 224kPa 6:16  
136.4 C 227kPa 6:46  
136.4 C 227kPa 6:46  
VENTING 6:46  
AIR DRYING 7:41  
CYCLE COMPLETE 8:07  
-----

Digital Signature #  
CBF276D2DB6DE1C2  
-----