

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2635

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2023 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
2:00 190.9 F 2.2  
4:00 189.5 F 6.2  
6:00 210.0 F 11.3  
8:00 220.1 F 6.5  
10:00 234.1 F 12.0  
12:00 234.5 F 9.5  
14:00 237.3 F 9.3  
16:00 248.0 F 14.9  
18:00 260.3 F 21.7  
20:00 271.3 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.4 F 29.8  
2:00 272.8 F 29.8  
3:00 272.2 F 29.4  
4:00 272.1 F 29.4

Min 271.0 F 28.9  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.9 F 0.0  
5:00 217.8 F 0.0  
10:00 225.1 F 0.0  
15:00 229.0 F 0.0  
20:00 229.8 F 0.0  
25:00 228.5 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 20:47  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

07/12/2023 10:29 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_