

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3728

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2023 03:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.0 F 0.0  
2:00 178.6 F 2.7  
4:00 199.9 F 7.7  
6:00 219.1 F 13.3  
8:00 230.1 F 8.2  
10:00 241.8 F 13.1  
12:00 240.8 F 11.2  
14:00 238.1 F 9.6  
16:00 252.1 F 16.3  
18:00 263.7 F 23.0  
20:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.1 F 28.6  
2:00 272.3 F 29.0  
3:00 271.6 F 28.2  
4:00 272.4 F 28.9

Min 271.2 F 28.2  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.7 F 0.0  
5:00 235.8 F 0.0  
10:00 224.5 F 0.0  
15:00 218.8 F 0.0  
20:00 206.8 F 0.0  
25:00 192.2 F 0.0  
30:00 192.4 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 20:03  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

09/06/2023 04:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_