
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000579
11:00 12/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:15
STERILIZING 4:22
136.0 C 228kPa 4:22
136.4 C 229kPa 4:52
136.3 C 231kPa 5:22
136.4 C 233kPa 5:52
136.4 C 233kPa 6:22
136.3 C 232kPa 6:52
136.4 C 232kPa 7:22
136.4 C 232kPa 7:52
136.4 C 232kPa 7:52
VENTING 7:52
AIR DRYING 8:46
CYCLE COMPLETE 23:46

Digital Signature #
11AA022863043E5B
