
Midmark M11 - v1.0.0
Total Cycles: 2675

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2022 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.6 F 0.0
2:00 201.4 F 3.9
4:00 209.6 F 8.8
6:00 225.6 F 7.5
8:00 234.5 F 10.1
10:00 237.1 F 7.7
12:00 244.6 F 12.1
14:00 245.5 F 12.9
16:00 258.6 F 19.9
18:00 270.1 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.2 F 28.7
2:00 271.2 F 28.1
3:00 272.2 F 28.8
4:00 271.4 F 28.1

Min 271.1 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 235.4 F 0.0
10:00 239.4 F 0.0
15:00 229.8 F 0.0
20:00 225.9 F 0.0
25:00 229.7 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

12/14/2022 01:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____