
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002226
9:32 06/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:31
STERILIZING 3:43
136.4 C 224kPa 3:43
136.1 C 223kPa 4:13
136.3 C 223kPa 4:43
135.8 C 220kPa 5:13
136.2 C 223kPa 5:43
136.3 C 224kPa 6:13
136.3 C 223kPa 6:43
136.4 C 225kPa 7:13
136.4 C 225kPa 7:13
VENTING 7:13
AIR DRYING 8:04
CYCLE COMPLETE 38:04

Digital Signature #
2DDAD4D8383F1223
