

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1319

Name: RITTER2  
Sterilizer ID: V2313228  
07/19/2022 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.0 F 0.0  
2:00 127.0 F 0.0  
4:00 152.4 F 0.3  
6:00 193.3 F 1.8  
8:00 220.6 F 6.1  
10:00 227.2 F 10.5  
12:00 228.9 F 7.3  
14:00 239.4 F 11.3  
16:00 235.5 F 8.9  
18:00 239.7 F 10.4  
20:00 247.4 F 14.2  
22:00 259.3 F 20.8  
24:00 270.0 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.4 F 29.3  
2:00 271.2 F 28.8  
3:00 272.5 F 29.4  
4:00 271.7 F 28.7

Min 271.2 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
5:00 214.9 F 0.0  
10:00 229.1 F 0.0  
15:00 226.2 F 0.0  
20:00 234.5 F 0.0  
25:00 233.7 F 0.0  
30:00 210.1 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 24:56  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:03:23  
CYCLE COMPLETE

07/19/2022 01:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_