

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005593  
13:31 24/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:21  
STERILIZING 3:17  
136.0 C 220kPa 3:17  
136.4 C 223kPa 3:47  
136.3 C 224kPa 4:17  
136.2 C 223kPa 4:47  
136.3 C 224kPa 5:17  
136.1 C 222kPa 5:47  
136.5 C 224kPa 6:17  
136.3 C 223kPa 6:47  
136.3 C 223kPa 6:47  
VENTING 6:47  
AIR DRYING 7:41  
CYCLE COMPLETE 28:28

-----

Digital Signature #  
25E553835485D844  
-----