
Midmark M11 - v1.0.5
Total Cycles: 2662

Name: RITTER2
Sterilizer ID: V2313228
07/18/2023 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.6 F 0.0
2:00 193.1 F 4.0
4:00 208.8 F 9.7
6:00 222.0 F 7.5
8:00 234.0 F 11.7
10:00 235.1 F 9.7
12:00 236.4 F 9.0
14:00 249.5 F 15.6
16:00 261.7 F 22.5
18:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.9 F 29.6
2:00 272.5 F 29.6
3:00 271.6 F 29.0
4:00 271.8 F 29.6

Min 271.3 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 227.2 F 0.0
10:00 230.4 F 0.0
15:00 227.4 F 0.0
20:00 224.7 F 0.0
25:00 224.3 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

07/18/2023 12:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____