
Midmark M11 - v1.0.5
Total Cycles: 658

Name: RITTER2
Sterilizer ID: V2313228
12/21/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 191.0 F 3.3
4:00 191.9 F 8.2
6:00 207.1 F 10.8
8:00 230.7 F 9.5
10:00 234.3 F 9.0
12:00 244.4 F 13.6
14:00 244.6 F 12.8
16:00 257.0 F 19.5
18:00 268.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.2 F 29.3
2:00 271.0 F 28.7
3:00 272.7 F 29.8
4:00 272.1 F 29.7

Min 271.0 F 28.6
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 233.5 F 0.0
10:00 229.9 F 0.0
15:00 229.0 F 0.0
20:00 231.0 F 0.0
25:00 230.0 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 2:23
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

12/21/2021 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____