

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1975

Name: RITTER2  
Sterilizer ID: V2313228  
01/30/2023 08:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.9 F 0.0  
2:00 81.5 F 0.0  
4:00 97.4 F 0.0  
6:00 124.4 F 0.0  
8:00 152.5 F 2.7  
10:00 175.0 F 5.5  
12:00 190.6 F 9.1  
14:00 210.9 F 13.0  
16:00 219.0 F 6.1  
18:00 230.4 F 9.9  
20:00 235.2 F 9.8  
22:00 240.5 F 11.4  
24:00 238.2 F 9.7  
26:00 250.8 F 15.9  
28:00 261.6 F 22.1  
30:00 271.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.5 F 29.2  
2:00 271.5 F 28.9  
3:00 272.4 F 29.2  
4:00 271.3 F 28.8

Min 271.3 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 220.5 F 0.0  
10:00 205.3 F 0.0  
15:00 189.9 F 0.0  
20:00 185.9 F 0.0  
25:00 182.4 F 0.0  
30:00 179.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:34  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:08:42  
CYCLE COMPLETE

01/30/2023 09:20 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_