
Midmark M11 - v1.0.5
Total Cycles: 2942

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.5 F 0.0
2:00 149.1 F 0.0
4:00 160.0 F 1.0
6:00 184.7 F 5.4
8:00 209.8 F 10.3
10:00 219.4 F 6.2
12:00 232.6 F 11.4
14:00 233.2 F 8.9
16:00 239.1 F 10.1
18:00 247.6 F 14.4
20:00 260.1 F 21.4
22:00 271.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.2 F 29.7
2:00 271.7 F 28.8
3:00 272.4 F 29.7
4:00 271.9 F 29.0

Min 271.4 F 28.8
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 199.0 F 0.0
10:00 203.6 F 0.0
15:00 201.3 F 0.0
20:00 200.9 F 0.0
25:00 204.5 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

09/28/2023 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____