
Midmark M11 - v1.0.5
Total Cycles: 2311

Name: RITTER2
Sterilizer ID: V2313228
04/28/2023 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 178.8 F 1.9
4:00 189.5 F 6.7
6:00 212.3 F 12.5
8:00 225.9 F 7.6
10:00 239.7 F 13.4
12:00 241.6 F 11.7
14:00 241.2 F 11.0
16:00 255.4 F 18.2
18:00 267.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.8
2:00 272.6 F 29.3
3:00 271.6 F 28.5
4:00 272.6 F 29.2

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 202.3 F 0.0
10:00 214.7 F 0.0
15:00 209.1 F 0.0
20:00 205.6 F 0.0
25:00 202.5 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

04/28/2023 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____