

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001813  
8:52 08/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:06  
STERILIZING 3:04  
136.4 C 223kPa 3:04  
136.4 C 226kPa 3:34  
136.4 C 224kPa 4:04  
136.4 C 225kPa 4:34  
136.4 C 226kPa 5:04  
136.2 C 223kPa 5:34  
136.4 C 225kPa 6:04  
136.4 C 225kPa 6:34  
136.4 C 225kPa 6:34  
VENTING 6:34  
AIR DRYING 7:26  
CYCLE COMPLETE 21:57

-----  
Digital Signature #  
CC421F19CC1B6B5C  
-----