
Midmark M11 - v1.0.0
Total Cycles: 4762

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2024 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 156.1 F 1.8
4:00 157.9 F 4.9
6:00 175.3 F 8.2
8:00 193.2 F 12.1
10:00 215.8 F 4.1
12:00 221.6 F 8.4
14:00 233.7 F 12.8
16:00 233.5 F 9.2
18:00 239.9 F 10.3
20:00 245.1 F 12.9
22:00 256.4 F 18.9
24:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.2 F 29.0
2:00 271.2 F 28.5
3:00 272.3 F 29.0
4:00 271.3 F 28.4

Min 271.2 F 28.4
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 213.4 F 0.0
10:00 210.9 F 0.0
15:00 214.1 F 0.0
20:00 239.6 F 0.0
25:00 237.4 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 25:38
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:18
CYCLE COMPLETE

07/24/2024 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____