

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1746

Name: RITTER2  
Sterilizer ID: V2313228  
11/17/2022 04:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
2:00 204.4 F 3.4  
4:00 216.5 F 9.4  
6:00 227.1 F 7.6  
8:00 238.3 F 12.3  
10:00 239.2 F 11.0  
12:00 239.6 F 10.7  
14:00 254.9 F 18.3  
16:00 267.3 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.0 F 29.2  
2:00 272.0 F 29.7  
3:00 271.8 F 29.0  
4:00 272.2 F 29.7  
  
Min 271.3 F 28.8  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 221.8 F 0.0  
10:00 225.0 F 0.0  
15:00 227.2 F 0.0  
20:00 225.9 F 0.0  
25:00 226.3 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 17:16  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:56:25  
CYCLE COMPLETE

11/17/2022 05:51 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_