
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000774
13:15 09/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:46
STERILIZING 4:55
136.4 C 220kPa 4:55
136.4 C 225kPa 5:25
136.4 C 224kPa 5:55
136.4 C 221kPa 6:25
136.2 C 223kPa 6:55
135.9 C 220kPa 7:25
136.4 C 224kPa 7:55
136.3 C 221kPa 8:25
136.3 C 221kPa 8:25
VENTING 8:25
AIR DRYING 9:21
CYCLE COMPLETE 9:23

Digital Signature #
FE3658270FECB5E4
