
Midmark M11 - v1.0.5
Total Cycles: 400

Name: RITTER2
Sterilizer ID: V2313228
09/28/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.7 F 0.0
2:00 110.7 F 0.9
4:00 131.0 F 3.5
6:00 147.8 F 6.9
8:00 178.3 F 10.9
10:00 203.6 F 3.3
12:00 220.8 F 7.9
14:00 234.5 F 13.0
16:00 236.5 F 10.4
18:00 235.3 F 8.8
20:00 249.8 F 15.6
22:00 261.5 F 22.4
24:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.4
2:00 272.0 F 29.7
3:00 272.0 F 29.1
4:00 272.6 F 29.9

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 210.1 F 0.0
10:00 213.4 F 0.0
15:00 224.7 F 0.0
20:00 220.9 F 0.0
25:00 219.9 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:24
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

09/28/2021 11:47 AM
ACCEPTED: X
REJECTED:
Operator: U01