
Midmark M11 - v1.0.5
Total Cycles: 127

Name: RITTER2
Sterilizer ID: V2313228
06/22/2021 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.4 F 0.0
2:00 134.7 F 1.5
4:00 147.4 F 4.0
6:00 175.0 F 7.4
8:00 194.1 F 11.7
10:00 213.2 F 5.3
12:00 223.3 F 9.6
14:00 232.8 F 10.3
16:00 238.4 F 11.5
18:00 238.0 F 10.0
20:00 250.0 F 16.0
22:00 260.6 F 22.1
24:00 270.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.7 F 30.0
2:00 271.3 F 29.4
3:00 272.6 F 30.2
4:00 271.9 F 29.6

Min 271.3 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 225.5 F 0.0
10:00 225.1 F 0.0
15:00 221.3 F 0.0
20:00 217.5 F 0.0
25:00 215.7 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:51
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

06/22/2021 02:20 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____