
Midmark M11 - v1.0.0
Total Cycles: 3208

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2023 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 75.4 F 0.0
4:00 82.8 F 0.0
6:00 110.2 F 0.0
8:00 146.5 F 0.2
10:00 173.6 F 3.3
12:00 193.7 F 6.4
14:00 215.6 F 9.5
16:00 229.9 F 13.4
18:00 228.8 F 7.2
20:00 238.7 F 11.4
22:00 234.9 F 7.8
24:00 245.8 F 13.5
26:00 243.5 F 11.8
28:00 255.5 F 18.0
30:00 266.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.7 F 28.0
2:00 271.8 F 28.1
3:00 272.1 F 28.3
4:00 272.3 F 28.5

Min 271.5 F 28.0
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 207.1 F 0.0
10:00 217.4 F 0.0
15:00 224.4 F 0.0
20:00 227.3 F 0.0
25:00 230.1 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:36
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

05/10/2023 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____