

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4167

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2024 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.5 F 0.0  
2:00 70.2 F 0.0  
4:00 77.3 F 0.0  
6:00 137.1 F 0.2  
8:00 145.6 F 0.6  
10:00 168.6 F 2.8  
12:00 191.3 F 6.0  
14:00 219.6 F 10.1  
16:00 230.2 F 8.1  
18:00 230.1 F 9.2  
20:00 239.1 F 11.1  
22:00 240.5 F 11.5  
24:00 238.2 F 10.0  
26:00 251.7 F 16.6  
28:00 263.2 F 23.4  
30:00 272.3 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.7 F 29.9  
2:00 272.3 F 29.1  
3:00 271.2 F 28.9  
4:00 272.4 F 29.9

Min 271.1 F 28.7  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 238.2 F 0.0  
10:00 213.4 F 0.0  
15:00 183.4 F 0.0  
20:00 182.0 F 0.0  
25:00 201.5 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:13  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:08:30  
CYCLE COMPLETE

11/29/2024 10:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_